


Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU030422\
 Data File : VU047323.D
 Acq On : 04 Mar 2022 08:59
 Operator : SY/MD
 Sample : BFB061
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFB061

Integration File: rteint.p

Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR030422WMA.M
 Title : TRACE VOA SFAM1.0
 Last Update : Fri Mar 04 23:20:39 2022



AutoFind: Scans 2890, 2891, 2892; Background Corrected with Scan 2876

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.3	31027	PASS
75	95	30	80	48.9	87819	PASS
95	95	100	100	100.0	179733	PASS
96	95	5	9	6.6	11915	PASS
173	174	0.00	2	0.9	1471	PASS
174	95	50	120	89.4	160704	PASS
175	174	5	9	7.2	11637	PASS
176	174	95	101	97.7	156949	PASS
177	176	5	9	6.4	10096	PASS

m/z	Abundance
35.90	1162.0
37.00	138.0
37.90	415.0
39.90	223.0
43.90	262.0

Instrument :
MSVOA_U
ClientSampleId :
BFB061

m/z	Abundance
35.90	2526.0
37.00	7923.0
38.00	7487.0
39.00	2713.0
39.90	305.0
43.00	100.0
43.90	1197.0
44.90	1260.0
45.90	143.0
47.00	1901.0
48.00	1026.0
49.00	7172.0
50.00	31872.0
51.00	10591.0
51.90	343.0
54.90	545.0
56.00	2532.0
57.00	4666.0
58.00	200.0
60.00	1593.0
61.00	8196.0
62.00	8086.0
63.00	6406.0
64.00	637.0
65.00	469.0
67.00	457.0
68.00	17832.0
69.00	17752.0
70.00	1377.0
72.00	925.0
73.00	7628.0
74.00	29816.0
75.00	88680.0
76.00	7970.0
77.00	1022.0
77.90	683.0
78.90	5507.0
79.90	1627.0
80.90	5782.0
81.90	1196.0
82.90	106.0
86.00	234.0
87.00	7584.0
87.90	7482.0
90.90	592.0
92.00	4976.0
93.00	7489.0
94.00	21176.0
95.00	179840.0
96.00	12382.0
97.00	419.0
102.80	102.0
103.90	783.0
104.80	323.0
105.90	713.0
106.90	121.0
109.90	103.0
110.80	154.0
111.90	122.0
112.80	102.0
114.80	141.0
115.90	602.0
116.90	1214.0
117.90	652.0
119.00	921.0
127.90	607.0
128.80	270.0
129.90	629.0
130.80	276.0
134.80	316.0
136.80	248.0
140.90	1564.0
141.80	234.0
142.90	1613.0
143.90	126.0
144.90	181.0
145.80	260.0
146.90	130.0
147.90	354.0
148.90	114.0
149.90	178.0
153.00	111.0
155.00	444.0
155.80	159.0
156.90	389.0
158.90	207.0
160.80	197.0
172.00	702.0
173.00	1459.0
173.90	157248.0
175.00	11462.0
175.90	153216.0

Instrument :
MSVOA_U
ClientSampleId :
BFB061

177.0010055.0

177.90320.0

Instrument :
MSVOA_U
ClientSampleId :
BFB061

m/z	Abundance
35.90	2521.0
37.00	8173.0
38.00	7596.0
39.00	2772.0
39.90	302.0
43.90	1189.0
45.00	1300.0
45.90	170.0
47.00	1927.0
48.00	1025.0
49.00	7326.0
50.00	31864.0
51.00	10640.0
51.90	348.0
55.00	536.0
56.00	2632.0
57.00	4883.0
58.00	216.0
60.00	1575.0
61.00	8568.0
62.00	8349.0
63.00	6449.0
64.00	632.0
64.90	411.0
67.00	510.0
68.00	18176.0
69.00	18320.0
70.00	1369.0
72.00	900.0
73.00	7633.0
74.00	30440.0
75.00	90424.0
76.00	7975.0
77.00	999.0
77.90	687.0
78.90	5581.0
79.90	1619.0
80.90	5981.0
81.90	1291.0
83.00	154.0
86.00	262.0
87.00	7806.0
87.90	7602.0
90.90	603.0
92.00	5141.0
93.00	7814.0
94.00	21648.0
95.00	185344.0
96.00	12369.0
97.00	396.0
103.90	847.0
104.80	291.0
105.90	815.0
106.90	148.0
109.90	126.0
110.90	182.0
111.90	112.0
112.70	121.0
114.90	149.0
115.90	611.0
116.90	1247.0
117.90	753.0
118.90	978.0
127.90	658.0
128.80	323.0
129.90	707.0
130.80	279.0
134.80	297.0
136.80	271.0
139.90	105.0
140.90	1678.0
141.90	232.0
142.90	1689.0
143.90	128.0
144.90	192.0
145.80	258.0
146.90	146.0
147.90	428.0
148.80	114.0
149.90	190.0
153.00	119.0
155.00	434.0
155.80	148.0
156.90	387.0
158.90	258.0
160.80	232.0
171.90	732.0
173.00	1524.0
173.90	165952.0
175.00	12013.0
175.90	162304.0
176.90	10409.0

Instrument :
MSVOA_U
ClientSampleId :
BFB061

177.90

294.0

Instrument :
MSVOA_U
ClientSampleId :
BFB061

m/z	Abundance
35.90	2405.0
37.00	7406.0
38.00	6847.0
39.00	2491.0
39.90	307.0
43.90	1080.0
45.00	1267.0
45.90	165.0
47.00	1745.0
48.00	927.0
49.00	6679.0
50.00	29344.0
51.00	9506.0
52.00	309.0
55.00	531.0
56.00	2424.0
57.00	4526.0
58.00	195.0
60.00	1449.0
61.00	7937.0
62.00	7831.0
63.00	5912.0
63.90	567.0
64.90	295.0
67.00	465.0
68.00	16864.0
69.00	17416.0
70.00	1247.0
71.90	774.0
73.00	7014.0
74.00	28336.0
75.00	84352.0
76.00	7192.0
77.00	915.0
77.80	664.0
78.90	5096.0
79.90	1480.0
80.90	5567.0
81.90	1272.0
83.00	162.0
86.10	264.0
87.00	7307.0
87.90	7061.0
90.90	513.0
92.00	4844.0
93.00	7399.0
94.00	20192.0
95.00	174016.0
96.00	10995.0
97.00	331.0
103.90	804.0
104.90	248.0
105.90	838.0
106.90	148.0
109.80	130.0
110.90	196.0
112.70	119.0
114.90	162.0
115.90	593.0
116.90	1148.0
117.90	699.0
118.90	959.0
127.80	658.0
128.90	328.0
129.90	679.0
130.80	243.0
134.80	257.0
136.80	274.0
139.80	114.0
140.90	1669.0
141.90	194.0
142.90	1685.0
143.90	118.0
144.90	188.0
145.80	216.0
146.90	146.0
147.90	455.0
149.90	175.0
152.90	107.0
154.90	382.0
155.90	130.0
156.90	312.0
158.90	276.0
160.80	239.0
171.90	680.0
173.00	1430.0
173.90	158912.0
175.00	11437.0
175.90	155328.0
176.90	9824.0
177.90	245.0

Instrument :
MSVOA_U
ClientSampleId :
BFB061