

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU030722\
 Data File : VU047369.D
 Acq On : 07 Mar 2022 15:55
 Operator : SY/MD
 Sample : N1810-01DL 5X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GBD87DL

Quant Time: Mar 08 03:01:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR030422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Mar 04 23:20:39 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	158086	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	149538	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	78035	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	55426	4.940	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	98.800%
7) Chloroethane-d5	1.919	69	53348	5.699	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	114.000%
11) 1,1-Dichloroethene-d2	2.571	65	23833	5.030	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	100.600%
20) 2-Butanone-d5	4.639	46	186203	50.550	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	101.100%
24) Chloroform-d	5.067	84	123674	5.966	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	119.400%
26) 1,2-Dichloroethane-d4	5.706	65	68240	5.865	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	117.400%
32) Benzene-d6	5.729	84	241954	5.745	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	114.800%
36) 1,2-Dichloropropane-d6	6.693	67	79307	6.046	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	121.000%
41) Toluene-d8	7.899	98	217626	5.473	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	109.400%
43) trans-1,3-Dichloroprop...	8.182	79	32153	5.578	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	111.600%
46) 2-Hexanone-d5	8.636	63	186311	56.792	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	113.580%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	74648	5.586	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	111.800%
66) 1,2-Dichlorobenzene-d4	12.195	152	98704	6.455	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	129.000%#
Target Compounds						
					Qvalue	
22) cis-1,2-Dichloroethene	4.671	96	1974	0.215	ug/L	90
34) Trichloroethene	6.546	95	78065	8.801	ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU030722\
Data File : VU047369.D
Acq On : 07 Mar 2022 15:55
Operator : SY/MD
Sample : N1810-01DL 5X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GBD87DL

Quant Time: Mar 08 03:01:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR030422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Mar 04 23:20:39 2022
Response via : Initial Calibration

