

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU030822\
 Data File : VU047396.D
 Acq On : 08 Mar 2022 17:22
 Operator : SY/MD
 Sample : N1840-01DL 5X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFY20DL

Quant Time: Mar 09 03:05:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR030422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Mar 09 02:01:55 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	210138	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	199279	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	110493	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	31370	2.103	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	42.000%	
7) Chloroethane-d5	1.919	69	34598	2.781	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	55.600%#	
11) 1,1-Dichloroethene-d2	2.571	65	13973	2.218	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	44.400%#	
20) 2-Butanone-d5	4.639	46	166776	34.061	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	68.120%	
24) Chloroform-d	5.066	84	90117	3.270	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	65.400%#	
26) 1,2-Dichloroethane-d4	5.706	65	53662	3.470	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	69.400%#	
32) Benzene-d6	5.729	84	165247	2.944	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	58.800%#	
36) 1,2-Dichloropropane-d6	6.693	67	60125	3.440	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	68.800%	
41) Toluene-d8	7.899	98	133791	2.525	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	50.400%#	
43) trans-1,3-Dichloroprop...	8.182	79	23901	3.112	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	62.200%	
46) 2-Hexanone-d5	8.635	63	172538	39.466	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	78.940%	
56) 1,1,2,2-Tetrachloroeth...	10.758	84	63567	3.569	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	71.400%	
66) 1,2-Dichlorobenzene-d4	12.195	152	71025	3.280	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	65.600%#	
Target Compounds						Qvalue
12) 1,1-Dichloroethene	2.584	96	4169	0.405	ug/L #	1
17) Methyl tert-butyl Ether	3.369	73	47672	1.728	ug/L	98
19) 1,1-Dichloroethane	3.883	63	2546	0.137	ug/L	96
22) cis-1,2-Dichloroethene	4.671	96	44169	3.627	ug/L	93
29) 1,1,1-Trichloroethane	5.324	97	2134	0.128	ug/L #	86
34) Trichloroethene	6.542	95	28640	2.423	ug/L	95
35) Methylcyclohexane	6.693	83	14124	0.779	ug/L #	25
37) 1,2-Dichloropropane	6.693	63	6349	0.569	ug/L #	88
47) Tetrachloroethene	8.555	164	36770	3.652	ug/L	91
49) Dibromochloromethane	8.555	129	33571	2.958	ug/L #	10
61) 1,2,3-Trichloropropane	11.812	75	11596	1.264	ug/L #	67


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU030822\
Data File : VU047396.D
Acq On : 08 Mar 2022 17:22
Operator : SY/MD
Sample : N1840-01DL 5X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFY20DL

Quant Time: Mar 09 03:05:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR030422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Mar 09 02:01:55 2022
Response via : Initial Calibration









#12
1,1-Dichloroethene
Concen: 0.405 ug/L
RT: 2.584 min Scan# 388
Delta R.T. -0.003 min
Lab File: VU047396.D
Acq: 08 Mar 2022 17:22




Instrument :
MSVOA_U
ClientSampleId :
BFY20DL


Tgt Ion: 96 Resp: 4169
Ion Ratio Lower Upper
96 100
61 210.7 124.7 231.5
63 634.1 105.2 195.4#



#17
Methyl tert-butyl Ether
Concen: 1.728 ug/L
RT: 3.369 min Scan# 631
Delta R.T. -0.000 min
Lab File: VU047396.D
Acq: 08 Mar 2022 17:22

Tgt Ion: 73 Resp: 47672
Ion Ratio Lower Upper
73 100
43 19.5 14.8 22.2
57 21.2 17.4 26.0





#19

1,1-Dichloroethane

Concen: 0.137 ug/L

RT: 3.883 min Scan# 791

Delta R.T. 0.006 min

Lab File: VU047396.D

Acq: 08 Mar 2022 17:22

Instrument :

MSVOA_U

ClientSampleId :

BFY20DL





Tgt Ion: 63 Resp: 2546

Ion Ratio Lower Upper

63 100

65 29.1 22.3 41.5

83 12.7 9.6 17.8



#22

cis-1,2-Dichloroethene

Concen: 3.627 ug/L

RT: 4.671 min Scan# 1036

Delta R.T. -0.000 min

Lab File: VU047396.D

Acq: 08 Mar 2022 17:22




Tgt Ion: 96 Resp: 44169


Ion Ratio Lower Upper

96 100

61 117.2 88.5 164.3

98 67.0 44.4 82.5









#29
1,1,1-Trichloroethane
Concen: 0.128 ug/L
RT: 5.324 min Scan# 11
Delta R.T. 0.003 min
Lab File: VU047396.D
Acq: 08 Mar 2022 17:22




Instrument :
MSVOA_U
ClientSampleId :
BFY20DL


Tgt Ion:	97	Resp:	2134
Ion	Ratio	Lower	Upper
97	100		
99	54.6	51.3	76.9
61	35.0	36.1	54.1#



#34
Trichloroethene
Concen: 2.423 ug/L
RT: 6.542 min Scan# 1618
Delta R.T. -0.003 min
Lab File: VU047396.D
Acq: 08 Mar 2022 17:22

Tgt Ion:	95	Resp:	28640
Ion	Ratio	Lower	Upper
95	100		
97	64.7	46.5	86.5
132	98.3	65.1	120.9
130	102.9	67.3	124.9









#35
Methylcyclohexane
Concen: 0.779 ug/L
RT: 6.693 min Scan# 1665
Delta R.T. -0.074 min
Lab File: VU047396.D
Acq: 08 Mar 2022 17:22




Instrument :
MSVOA_U
ClientSampleId :
BFY20DL


Tgt Ion: 83 Resp: 14124
Ion Ratio Lower Upper
83 100
55 10.0 61.9 92.9#
98 0.4 39.0 58.6#



#37
1,2-Dichloropropane
Concen: 0.569 ug/L
RT: 6.693 min Scan# 1665
Delta R.T. -0.100 min
Lab File: VU047396.D
Acq: 08 Mar 2022 17:22

Tgt Ion: 63 Resp: 6349
Ion Ratio Lower Upper
63 100
112 0.0 3.2 4.8#









#47
Tetrachloroethene
Concen: 3.652 ug/L
RT: 8.555 min Scan# 211
Delta R.T. -0.000 min
Lab File: VU047396.D
Acq: 08 Mar 2022 17:22




Instrument :
MSVOA_U
ClientSampleId :
BFY20DL


Tgt Ion:164 Resp: 36770
Ion Ratio Lower Upper
164 100
129 89.3 71.8 133.4
131 90.0 68.5 127.3
166 125.5 94.1 174.8



#49
Dibromochloromethane
Concen: 2.958 ug/L
RT: 8.555 min Scan# 2244
Delta R.T. -0.257 min
Lab File: VU047396.D
Acq: 08 Mar 2022 17:22

Tgt Ion:129 Resp: 33571
Ion Ratio Lower Upper
129 100
127 0.0 55.4 102.8#





#61

1,2,3-Trichloropropane

Concen: 1.264 ug/L

RT: 11.812 min Scan# 31

Delta R.T. 0.987 min

Lab File: VU047396.D


Acq: 08 Mar 2022 17:22

Instrument :

MSVOA_U

ClientSampleId :

BFY20DL



Tgt Ion: 75 Resp: 11596

Ion Ratio Lower Upper

75 100

110 2.1 29.9 44.9#

77 32.9 25.5 38.3

