

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU030922\
 Data File : VU047438.D
 Acq On : 10 Mar 2022 06:05
 Operator : SY/MD
 Sample : N1857-08
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 ETNH7

Quant Time: Mar 10 06:34:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR030922WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Mar 10 03:57:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	159663	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	154329	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	81503	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	39008	4.505	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	90.000%	
7) Chloroethane-d5	1.919	69	30099	4.735	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	94.600%	
11) 1,1-Dichloroethene-d2	2.572	65	17781	4.649	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	93.000%	
20) 2-Butanone-d5	4.649	46	70976	46.037	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	92.080%	
24) Chloroform-d	5.067	84	62868	4.557	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	91.200%	
26) 1,2-Dichloroethane-d4	5.706	65	32949	4.717	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	94.400%	
32) Benzene-d6	5.732	84	136365	4.697	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	94.000%	
36) 1,2-Dichloropropane-d6	6.694	67	40267	4.689	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	93.800%	
41) Toluene-d8	7.899	98	128643	4.508	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	90.200%	
43) trans-1,3-Dichloroprop...	8.182	79	15041	4.125	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	82.600%	
46) 2-Hexanone-d5	8.639	63	73934	46.030	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	92.060%	
56) 1,1,2,2-Tetrachloroeth...	10.758	84	34252	4.577	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	91.600%	
66) 1,2-Dichlorobenzene-d4	12.195	152	51315	4.966	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	99.400%	
Target Compounds						
5) Vinyl chloride	1.607	62	5393	0.459	ug/L	84
8) Chloroethane	1.945	64	1229	0.172	ug/L	87
18) trans-1,2-Dichloroethene	3.359	96	885	0.088	ug/L	75
22) cis-1,2-Dichloroethene	4.674	96	19927	1.792	ug/L	97
33) Benzene	5.780	78	6479	0.155	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU030922\
Data File : VU047438.D
Acq On : 10 Mar 2022 06:05
Operator : SY/MD
Sample : N1857-08
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
ETNH7

Quant Time: Mar 10 06:34:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR030922WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Mar 10 03:57:44 2022
Response via : Initial Calibration

