

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU030922\
 Data File : VU047432.D
 Acq On : 10 Mar 2022 03:44
 Operator : SY/MD
 Sample : N1857-02
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 ETNH0

Quant Time: Mar 10 06:33:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR030922WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Mar 10 03:57:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	160984	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	154054	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	83529	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	40289	4.614	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	92.200%
7) Chloroethane-d5	1.919	69	30485	4.756	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	95.200%
11) 1,1-Dichloroethene-d2	2.572	65	18169	4.711	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	94.200%
20) 2-Butanone-d5	4.645	46	73351	47.187	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	94.380%
24) Chloroform-d	5.070	84	64186	4.615	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.200%
26) 1,2-Dichloroethane-d4	5.707	65	33954	4.821	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.400%
32) Benzene-d6	5.732	84	139368	4.809	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.200%
36) 1,2-Dichloropropane-d6	6.694	67	41359	4.825	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.400%
41) Toluene-d8	7.899	98	132490	4.652	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.000%
43) trans-1,3-Dichloroprop...	8.182	79	16751	4.602	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.000%
46) 2-Hexanone-d5	8.639	63	73449	45.810	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	91.620%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	34062	4.559	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.200%
66) 1,2-Dichlorobenzene-d4	12.195	152	51783	4.890	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU030922\
Data File : VU047432.D
Acq On : 10 Mar 2022 03:44
Operator : SY/MD
Sample : N1857-02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
ETNH0

Quant Time: Mar 10 06:33:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR030922WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Mar 10 03:57:44 2022
Response via : Initial Calibration

