

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031022\
 Data File : VU047457.D
 Acq On : 10 Mar 2022 14:53
 Operator : SY/MD
 Sample : N1810-01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GBD87

Quant Time: Mar 11 00:54:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR030922WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Mar 11 00:52:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	164394	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	159540	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	83789	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	40758	4.571	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	91.400%
7) Chloroethane-d5	1.919	69	30973	4.732	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.600%
11) 1,1-Dichloroethene-d2	2.572	65	18337	4.656	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	93.200%
20) 2-Butanone-d5	4.645	46	80958	51.000	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	102.000%
24) Chloroform-d	5.067	84	67160	4.728	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.600%
26) 1,2-Dichloroethane-d4	5.706	65	35692	4.962	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.200%
32) Benzene-d6	5.732	84	145463	4.847	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.000%
36) 1,2-Dichloropropane-d6	6.694	67	43461	4.895	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.000%
41) Toluene-d8	7.899	98	134887	4.573	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.400%
43) trans-1,3-Dichloroprop...	8.182	79	17325	4.596	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.000%
46) 2-Hexanone-d5	8.639	63	77956	46.949	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.900%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	36025	4.656	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.200%
66) 1,2-Dichlorobenzene-d4	12.195	152	52322	4.925	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.600%
Target Compounds						Qvalue
12) 1,1-Dichloroethene	2.581	96	1449	0.150	ug/L	# 1
22) cis-1,2-Dichloroethene	4.674	96	8454	0.738	ug/L	97
34) Trichloroethene	6.546	95	351706	30.535	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031022\
Data File : VU047457.D
Acq On : 10 Mar 2022 14:53
Operator : SY/MD
Sample : N1810-01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GBD87

Quant Time: Mar 11 00:54:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR030922WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Mar 11 00:52:41 2022
Response via : Initial Calibration

