

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031022\
 Data File : VU047458.D
 Acq On : 10 Mar 2022 15:16
 Operator : SY/MD
 Sample : N1810-23
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Mar 11 00:55:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR030922WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Mar 11 00:52:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	150662	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	143766	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	76967	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	39987	4.894	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	97.800%	
7) Chloroethane-d5	1.919	69	30693	5.117	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	102.400%	
11) 1,1-Dichloroethene-d2	2.572	65	17917	4.964	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	99.200%	
20) 2-Butanone-d5	4.646	46	90312	62.078	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	124.160%	
24) Chloroform-d	5.067	84	65170	5.006	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	100.200%	
26) 1,2-Dichloroethane-d4	5.707	65	33773	5.123	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	102.400%	
32) Benzene-d6	5.732	84	138680	5.128	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	102.600%	
36) 1,2-Dichloropropane-d6	6.694	67	42129	5.266	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	105.400%	
41) Toluene-d8	7.899	98	130251	4.900	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	98.000%	
43) trans-1,3-Dichloroprop...	8.182	79	16631	4.896	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	98.000%	
46) 2-Hexanone-d5	8.639	63	74080	49.510	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	99.020%	
56) 1,1,2,2-Tetrachloroeth...	10.758	84	34650	4.970	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	99.400%	
66) 1,2-Dichlorobenzene-d4	12.195	152	51210	5.248	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	105.000%	
Target Compounds						Qvalue
3) Chloromethane	1.523	50	1372	0.122	ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031022\
Data File : VU047458.D
Acq On : 10 Mar 2022 15:16
Operator : SY/MD
Sample : N1810-23
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Mar 11 00:55:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR030922WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Mar 11 00:52:41 2022
Response via : Initial Calibration

