

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031022\
 Data File : VU047472.D
 Acq On : 10 Mar 2022 20:47
 Operator : SY/MD
 Sample : N1861-06DL 5X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 YATMODL

Quant Time: Mar 11 00:58:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR030922WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Mar 11 00:52:41 2022
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene		6.253	114	126361	5.000	ug/L	0.00
28) Chlorobenzene-d5		9.420	117	119669	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4		11.812	152	61363	5.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3		1.600	65	35157	5.130	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	102.600%	
7) Chloroethane-d5		1.919	69	26701	5.307	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	106.200%	
11) 1,1-Dichloroethene-d2		2.571	65	15888	5.249	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	105.000%	
20) 2-Butanone-d5		4.648	46	62434	51.169	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	102.340%	
24) Chloroform-d		5.066	84	57643	5.280	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.600%	
26) 1,2-Dichloroethane-d4		5.706	65	30662	5.546	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	111.000%	
32) Benzene-d6		5.732	84	121496	5.397	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.000%	
36) 1,2-Dichloropropane-d6		6.693	67	36182	5.433	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	108.600%	
41) Toluene-d8		7.899	98	109927	4.968	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.400%	
43) trans-1,3-Dichloroprop...		8.182	79	14246	5.039	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	100.800%	
46) 2-Hexanone-d5		8.639	63	65176	52.330	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	104.660%	
56) 1,1,2,2-Tetrachloroeth...		10.758	84	30549	5.264	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	105.200%	
66) 1,2-Dichlorobenzene-d4		12.195	152	43911	5.644	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	112.800%	
Target Compounds							
12) 1,1-Dichloroethene		2.584	96	1051	0.142	ug/L	Qvalue # 1
13) Acetone		2.668	43	9220	8.224	ug/L	58
34) Trichloroethene		6.545	95	56786	6.573	ug/L	99
47) Tetrachloroethene		8.555	164	14219	1.940	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031022\
Data File : VU047472.D
Acq On : 10 Mar 2022 20:47
Operator : SY/MD
Sample : N1861-06DL 5X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
YATMODL

Quant Time: Mar 11 00:58:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR030922WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Mar 11 00:52:41 2022
Response via : Initial Calibration

