

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031022\
 Data File : VU047454.D
 Acq On : 10 Mar 2022 13:39
 Operator : SY/MD
 Sample : VU0310WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK076

Quant Time: Mar 11 00:54:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR030922WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Mar 11 00:52:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	146931	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	140430	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	74580	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	38038	4.773	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	95.400%	
7) Chloroethane-d5	1.919	69	28165	4.814	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	96.200%	
11) 1,1-Dichloroethene-d2	2.575	65	17172	4.879	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	97.600%	
20) 2-Butanone-d5	4.645	46	77210	54.420	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	108.840%	
24) Chloroform-d	5.067	84	61456	4.841	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	96.800%	
26) 1,2-Dichloroethane-d4	5.707	65	32311	5.026	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	100.600%	
32) Benzene-d6	5.732	84	131208	4.967	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	99.400%	
36) 1,2-Dichloropropane-d6	6.694	67	39351	5.036	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	100.800%	
41) Toluene-d8	7.899	98	122186	4.706	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	94.200%	
43) trans-1,3-Dichloroprop...	8.182	79	15664	4.721	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	94.400%	
46) 2-Hexanone-d5	8.639	63	69823	47.773	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	95.540%	
56) 1,1,2,2-Tetrachloroeth...	10.758	84	31937	4.690	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	93.800%	
66) 1,2-Dichlorobenzene-d4	12.195	152	47061	4.977	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	99.600%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031022\
Data File : VU047454.D
Acq On : 10 Mar 2022 13:39
Operator : SY/MD
Sample : VU0310WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK076

Quant Time: Mar 11 00:54:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR030922WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Mar 11 00:52:41 2022
Response via : Initial Calibration

