

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031022\
 Data File : VU047456.D
 Acq On : 10 Mar 2022 14:29
 Operator : SY/MD
 Sample : N1773-05
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Mar 11 00:54:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR030922WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Mar 11 00:52:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	157790	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	153693	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	80401	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	40100	4.686	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	93.800%
7) Chloroethane-d5	1.919	69	30834	4.908	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	98.200%
11) 1,1-Dichloroethene-d2	2.572	65	18528	4.902	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	98.000%
20) 2-Butanone-d5	4.649	46	62618	41.098	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	82.200%
24) Chloroform-d	5.067	84	66855	4.904	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.000%
26) 1,2-Dichloroethane-d4	5.706	65	34955	5.063	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.200%
32) Benzene-d6	5.732	84	144081	4.984	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.600%
36) 1,2-Dichloropropane-d6	6.694	67	42398	4.957	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.200%
41) Toluene-d8	7.899	98	133505	4.698	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.000%
43) trans-1,3-Dichloroprop...	8.182	79	17547	4.832	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	96.600%
46) 2-Hexanone-d5	8.639	63	76738	47.974	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	95.940%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	35014	4.698	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.000%
66) 1,2-Dichlorobenzene-d4	12.195	152	51682	5.070	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031022\
Data File : VU047456.D
Acq On : 10 Mar 2022 14:29
Operator : SY/MD
Sample : N1773-05
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Mar 11 00:54:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR030922WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Mar 11 00:52:41 2022
Response via : Initial Calibration

