

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU031122\  
 Data File : VU047508.D  
 Acq On : 11 Mar 2022 21:54  
 Operator : SY/MD  
 Sample : N1909-04  
 Misc : 25.0mL/MSVOA\_U/WATER  
 ALS Vial : 33 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampleId :  
 YATN7

Quant Time: Mar 12 04:22:41 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\SFAMUTR030922WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Sat Mar 12 03:24:52 2022  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	124740	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	121245	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	63311	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	28275	4.179	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	83.600%	
7) Chloroethane-d5	1.919	69	22675	4.566	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	91.400%	
11) 1,1-Dichloroethene-d2	2.575	65	12831	4.294	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	85.800%	
20) 2-Butanone-d5	4.642	46	65647	54.501	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	109.000%	
24) Chloroform-d	5.070	84	51149	4.746	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	95.000%	
26) 1,2-Dichloroethane-d4	5.707	65	26149	4.791	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	95.800%	
32) Benzene-d6	5.729	84	102188	4.481	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	89.600%	
36) 1,2-Dichloropropane-d6	6.694	67	31091	4.608	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	92.200%	
41) Toluene-d8	7.899	98	92425	4.123	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	82.400%	
43) trans-1,3-Dichloroprop...	8.182	79	12510	4.367	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	87.400%	
46) 2-Hexanone-d5	8.636	63	56919	45.107	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	90.220%	
56) 1,1,2,2-Tetrachloroeth...	10.758	84	28388	4.828	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	96.600%	
66) 1,2-Dichlorobenzene-d4	12.192	152	39062	4.866	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	97.400%	
Target Compounds						
13) Acetone	2.658	43	63186	57.095	ug/L	95
25) Chloroform	5.096	83	11233	0.778	ug/L	100
34) Trichloroethene	6.546	95	24368	2.784	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU031122\  
 Data File : VU047508.D  
 Acq On : 11 Mar 2022 21:54  
 Operator : SY/MD  
 Sample : N1909-04  
 Misc : 25.0mL/MSVOA\_U/WATER  
 ALS Vial : 33 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampleId :  
 YATN7

Quant Time: Mar 12 04:22:41 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\SFAMUTR030922WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Sat Mar 12 03:24:52 2022  
 Response via : Initial Calibration

