

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031124\
 Data File : VU058035.D
 Acq On : 11 Mar 2024 12:07
 Operator : MD/SY
 Sample : P1636-23
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
ClientSampleId :
 DCNQ7

Quant Time: Mar 12 03:07:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR022924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Mar 12 03:05:49 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	146314	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	141367	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	63217	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	39732	3.774	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	75.400%	
7) Chloroethane-d5	1.907	69	36813	4.255	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	85.000%	
11) 1,1-Dichloroethene-d2	2.560	65	18691	4.103	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	82.000%	
20) 2-Butanone-d5	4.611	46	93402	53.849	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	107.700%	
24) Chloroform-d	5.052	84	90879	4.564	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	91.200%	
26) 1,2-Dichloroethane-d4	5.695	65	44172	4.896	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	98.000%	
32) Benzene-d6	5.718	84	179209	4.350	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	87.000%	
36) 1,2-Dichloropropane-d6	6.682	67	57601	4.599	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	92.000%	
41) Toluene-d8	7.891	98	153193	4.104	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	82.000%	
43) trans-1,3-Dichloroprop...	8.174	79	17604	4.088	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	81.800%	
46) 2-Hexanone-d5	8.627	63	73949	51.141	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	102.280%	
56) 1,1,2,2-Tetrachloroeth...	10.749	84	36931	4.809	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	96.200%	
66) 1,2-Dichlorobenzene-d4	12.190	152	51370	4.934	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	98.600%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031124\
 Data File : VU058035.D
 Acq On : 11 Mar 2024 12:07
 Operator : MD/SY
 Sample : P1636-23
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DCNQ7

Quant Time: Mar 12 03:07:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR022924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Mar 12 03:05:49 2024
 Response via : Initial Calibration

