

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031424\
 Data File : VU058094.D
 Acq On : 14 Mar 2024 11:58
 Operator : MD/SY
 Sample : P1741-12
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Mar 15 01:20:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR022924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Mar 14 04:37:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	141371	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	137097	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	60264	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	52567	5.168	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	103.400%
7) Chloroethane-d5	1.907	69	43699	5.227	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	104.600%
11) 1,1-Dichloroethene-d2	2.560	65	23255	5.283	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	105.600%
20) 2-Butanone-d5	4.615	46	83975	50.107	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	100.220%
24) Chloroform-d	5.052	84	96370	5.009	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.200%
26) 1,2-Dichloroethane-d4	5.695	65	45556	5.226	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.600%
32) Benzene-d6	5.718	84	197195	4.936	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.800%
36) 1,2-Dichloropropane-d6	6.682	67	61438	5.058	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.200%
41) Toluene-d8	7.891	98	167585	4.629	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.600%
43) trans-1,3-Dichloroprop...	8.174	79	17543	4.200	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.000%
46) 2-Hexanone-d5	8.627	63	67974	48.473	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	96.940%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	34935	4.691	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.800%
66) 1,2-Dichlorobenzene-d4	12.187	152	51032	5.142	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031424\
Data File : VU058094.D
Acq On : 14 Mar 2024 11:58
Operator : MD/SY
Sample : P1741-12
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Mar 15 01:20:03 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR022924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Mar 14 04:37:22 2024
Response via : Initial Calibration

