

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031424\
 Data File : VU058096.D
 Acq On : 14 Mar 2024 12:45
 Operator : MD/SY
 Sample : P1729-02DL 10X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AJ1DL

Quant Time: Mar 15 01:20:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR022924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Mar 14 04:37:22 2024
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene		6.242	114	129867	5.000	ug/L	0.00
28) Chlorobenzene-d5		9.412	117	123694	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4		11.807	152	55355	5.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3		1.592	65	53025	5.675	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	113.400%	
7) Chloroethane-d5		1.907	69	44338	5.773	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	115.400%	
11) 1,1-Dichloroethene-d2		2.560	65	23522	5.817	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	116.400%	
20) 2-Butanone-d5		4.618	46	91470	59.414	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	118.820%	
24) Chloroform-d		5.052	84	96122	5.439	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.800%	
26) 1,2-Dichloroethane-d4		5.695	65	45022	5.623	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	112.400%	
32) Benzene-d6		5.717	84	195442	5.422	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.400%	
36) 1,2-Dichloropropane-d6		6.682	67	61229	5.587	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	111.800%	
41) Toluene-d8		7.891	98	170908	5.233	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.600%	
43) trans-1,3-Dichloroprop...		8.177	79	17864	4.741	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	94.800%	
46) 2-Hexanone-d5		8.627	63	73949	58.448	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	116.900%	
56) 1,1,2,2-Tetrachloroeth...		10.749	84	36239	5.393	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	107.800%	
66) 1,2-Dichlorobenzene-d4		12.187	152	52319	5.739	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	114.800%	
Target Compounds							
						Qvalue	
12) 1,1-Dichloroethene		2.570	96	2491	0.318	ug/L	# 1
16) Methylene chloride		3.039	84	2009	0.211	ug/L	98
22) cis-1,2-Dichloroethene		4.660	96	13110	1.396	ug/L	94
34) Trichloroethene		6.534	95	92182	9.230	ug/L	99
47) Tetrachloroethene		8.550	164	3356	0.484	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031424\
Data File : VU058096.D
Acq On : 14 Mar 2024 12:45
Operator : MD/SY
Sample : P1729-02DL 10X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AJ1DL

Quant Time: Mar 15 01:20:18 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR022924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Mar 14 04:37:22 2024
Response via : Initial Calibration

