

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031424\
 Data File : VU058103.D
 Acq On : 14 Mar 2024 15:46
 Operator : MD/SY
 Sample : P1741-18
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0GJ3

Quant Time: Mar 15 01:21:13 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR022924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Mar 14 04:37:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)

Internal Standards					
1) 1,4-Difluorobenzene	6.242	114	129699	5.000 ug/L	0.00
28) Chlorobenzene-d5	9.412	117	121703	5.000 ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	56441	5.000 ug/L	0.00
System Monitoring Compounds					
4) Vinyl Chloride-d3	1.592	65	53101	5.690 ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 113.800%	
7) Chloroethane-d5	1.908	69	45649	5.952 ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 119.000%	
11) 1,1-Dichloroethene-d2	2.560	65	23461	5.810 ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 116.200%	
20) 2-Butanone-d5	4.612	46	85681	55.726 ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	= 111.460%	
24) Chloroform-d	5.052	84	99368	5.630 ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 112.600%	
26) 1,2-Dichloroethane-d4	5.695	65	46662	5.835 ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 116.800%	
32) Benzene-d6	5.721	84	200497	5.653 ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 113.000%	
36) 1,2-Dichloropropane-d6	6.682	67	62475	5.794 ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 115.800%	
41) Toluene-d8	7.891	98	170626	5.310 ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 106.200%	
43) trans-1,3-Dichloroprop...	8.177	79	17380	4.688 ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 93.800%	
46) 2-Hexanone-d5	8.627	63	72140	57.951 ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 115.900%	
56) 1,1,2,2-Tetrachloroeth...	10.750	84	38478	5.820 ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 116.400%	
66) 1,2-Dichlorobenzene-d4	12.187	152	53725	5.780 ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 115.600%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031424\
Data File : VU058103.D
Acq On : 14 Mar 2024 15:46
Operator : MD/SY
Sample : P1741-18
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0GJ3

Quant Time: Mar 15 01:21:13 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR022924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Mar 14 04:37:22 2024
Response via : Initial Calibration

