

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031424\
 Data File : VU058107.D
 Acq On : 14 Mar 2024 17:21
 Operator : MD/SY
 Sample : P1741-21
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0GK2

Quant Time: Mar 15 01:21:43 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR022924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Mar 14 04:37:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	127504	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	124758	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	56041	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	54268	5.916	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 118.400%		
7) Chloroethane-d5	1.907	69	46089	6.113	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 122.200%		
11) 1,1-Dichloroethene-d2	2.560	65	24245	6.107	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 122.200%		
20) 2-Butanone-d5	4.611	46	89638	59.303	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	= 118.600%		
24) Chloroform-d	5.052	84	101662	5.859	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 117.200%		
26) 1,2-Dichloroethane-d4	5.692	65	48217	6.133	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 122.600%		
32) Benzene-d6	5.717	84	203091	5.586	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 111.800%		
36) 1,2-Dichloropropane-d6	6.682	67	64049	5.795	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 115.800%		
41) Toluene-d8	7.891	98	172419	5.234	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 104.600%		
43) trans-1,3-Dichloroprop...	8.177	79	18169	4.780	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 95.600%		
46) 2-Hexanone-d5	8.627	63	75309	59.015	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 118.040%		
56) 1,1,2,2-Tetrachloroeth...	10.749	84	39031	5.759	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 115.200%		
66) 1,2-Dichlorobenzene-d4	12.190	152	54428	5.897	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 118.000%		
Target Compounds						
					Qvalue	
13) Acetone	2.618	43	8876	7.891	ug/L	97
25) Chloroform	5.078	83	21075	1.235	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031424\
Data File : VU058107.D
Acq On : 14 Mar 2024 17:21
Operator : MD/SY
Sample : P1741-21
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0GK2

Quant Time: Mar 15 01:21:43 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR022924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Mar 14 04:37:22 2024
Response via : Initial Calibration

