

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031424\
 Data File : VU058114.D
 Acq On : 14 Mar 2024 20:06
 Operator : MD/SY
 Sample : P1765-07
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0GM2

Quant Time: Mar 15 01:22:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR022924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Mar 14 04:37:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	129716	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	125118	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	55143	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	46073	4.937	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	98.800%
7) Chloroethane-d5	1.907	69	40447	5.273	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	105.400%
11) 1,1-Dichloroethene-d2	2.560	65	21723	5.379	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	107.600%
20) 2-Butanone-d5	4.611	46	79892	51.954	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.900%
24) Chloroform-d	5.052	84	88007	4.986	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.800%
26) 1,2-Dichloroethane-d4	5.695	65	43152	5.395	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	108.000%
32) Benzene-d6	5.721	84	179632	4.927	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.600%
36) 1,2-Dichloropropane-d6	6.682	67	56704	5.116	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.400%
41) Toluene-d8	7.894	98	154850	4.687	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.800%
43) trans-1,3-Dichloroprop...	8.177	79	15714	4.123	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	82.400%
46) 2-Hexanone-d5	8.627	63	65225	50.966	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	101.940%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	33282	4.896	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.000%
66) 1,2-Dichlorobenzene-d4	12.190	152	49109	5.408	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	108.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031424\
Data File : VU058114.D
Acq On : 14 Mar 2024 20:06
Operator : MD/SY
Sample : P1765-07
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0GM2

Quant Time: Mar 15 01:22:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR022924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Mar 14 04:37:22 2024
Response via : Initial Calibration

