

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031424\
 Data File : VU058116.D
 Acq On : 14 Mar 2024 20:53
 Operator : MD/SY
 Sample : P1765-09
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 COGM4

Quant Time: Mar 15 01:22:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR022924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Mar 14 04:37:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	130402	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	125714	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	54618	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	47379	5.050	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 101.000%		
7) Chloroethane-d5	1.907	69	41833	5.425	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 108.400%		
11) 1,1-Dichloroethene-d2	2.560	65	21528	5.302	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 106.000%		
20) 2-Butanone-d5	4.611	46	82846	53.592	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	= 107.180%		
24) Chloroform-d	5.052	84	92595	5.218	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 104.400%		
26) 1,2-Dichloroethane-d4	5.695	65	44394	5.521	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 110.400%		
32) Benzene-d6	5.721	84	189487	5.172	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 103.400%		
36) 1,2-Dichloropropane-d6	6.682	67	58215	5.227	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 104.600%		
41) Toluene-d8	7.891	98	161955	4.879	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 97.600%		
43) trans-1,3-Dichloroprop...	8.177	79	16660	4.350	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 87.000%		
46) 2-Hexanone-d5	8.627	63	69126	53.758	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 107.520%		
56) 1,1,2,2-Tetrachloroeth...	10.749	84	34903	5.111	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 102.200%		
66) 1,2-Dichlorobenzene-d4	12.190	152	50359	5.599	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 112.000%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031424\
Data File : VU058116.D
Acq On : 14 Mar 2024 20:53
Operator : MD/SY
Sample : P1765-09
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0GM4

Quant Time: Mar 15 01:22:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR022924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Mar 14 04:37:22 2024
Response via : Initial Calibration

