

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031424\
 Data File : VU058104.D
 Acq On : 14 Mar 2024 16:10
 Operator : MD/SY
 Sample : P1741-14
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0GK8

Quant Time: Mar 15 01:21:22 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR022924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Mar 14 04:37:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.241	114	129982	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	124362	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	55100	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	54697	5.849	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	117.000%
7) Chloroethane-d5	1.907	69	48196	6.270	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	125.400%
11) 1,1-Dichloroethene-d2	2.557	65	24272	5.998	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	120.000%
20) 2-Butanone-d5	4.611	46	90936	59.015	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	118.020%
24) Chloroform-d	5.052	84	103394	5.845	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	117.000%
26) 1,2-Dichloroethane-d4	5.695	65	49056	6.121	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	122.400%
32) Benzene-d6	5.721	84	209289	5.775	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	115.400%
36) 1,2-Dichloropropane-d6	6.682	67	65116	5.910	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	118.200%
41) Toluene-d8	7.891	98	180991	5.512	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	110.200%
43) trans-1,3-Dichloroprop...	8.177	79	18945	5.000	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	100.000%
46) 2-Hexanone-d5	8.627	63	75371	59.252	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	118.500%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	39715	5.878	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	117.600%
66) 1,2-Dichlorobenzene-d4	12.187	152	56521	6.229	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	124.600%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031424\
Data File : VU058104.D
Acq On : 14 Mar 2024 16:10
Operator : MD/SY
Sample : P1741-14
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0GK8

Quant Time: Mar 15 01:21:22 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR022924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Mar 14 04:37:22 2024
Response via : Initial Calibration

