

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031424\
 Data File : VU058105.D
 Acq On : 14 Mar 2024 16:34
 Operator : MD/SY
 Sample : P1741-19
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0GJ6

Quant Time: Mar 15 01:21:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR022924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Mar 14 04:37:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	127699	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	125887	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	55617	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	50318	5.477	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	109.600%
7) Chloroethane-d5	1.907	69	43245	5.727	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	114.600%
11) 1,1-Dichloroethene-d2	2.560	65	23022	5.790	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	115.800%
20) 2-Butanone-d5	4.611	46	85517	56.490	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	112.980%
24) Chloroform-d	5.052	84	94561	5.442	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.800%
26) 1,2-Dichloroethane-d4	5.695	65	45978	5.840	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	116.800%
32) Benzene-d6	5.721	84	193832	5.284	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.600%
36) 1,2-Dichloropropane-d6	6.682	67	60113	5.390	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	107.800%
41) Toluene-d8	7.891	98	166624	5.013	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.200%
43) trans-1,3-Dichloroprop...	8.177	79	17446	4.549	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.000%
46) 2-Hexanone-d5	8.627	63	70649	54.867	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	109.740%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	37536	5.489	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	109.800%
66) 1,2-Dichlorobenzene-d4	12.190	152	51423	5.614	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	112.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031424\
Data File : VU058105.D
Acq On : 14 Mar 2024 16:34
Operator : MD/SY
Sample : P1741-19
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0GJ6

Quant Time: Mar 15 01:21:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR022924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Mar 14 04:37:22 2024
Response via : Initial Calibration

