

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031522\
 Data File : VU047571.D
 Acq On : 15 Mar 2022 12:47
 Operator : SY/MD
 Sample : N1878-09
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Mar 16 06:45:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR031422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Mar 16 06:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	132501	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	131123	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	66132	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	34461	3.977	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	79.600%
7) Chloroethane-d5	1.919	69	35907	4.566	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	91.400%
11) 1,1-Dichloroethene-d2	2.571	65	15276	3.759	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	75.200%
20) 2-Butanone-d5	4.645	46	134159	53.807	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	107.620%
24) Chloroform-d	5.066	84	84977	4.466	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.400%
26) 1,2-Dichloroethane-d4	5.706	65	48451	4.699	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.000%
32) Benzene-d6	5.732	84	160751	4.211	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.200%
36) 1,2-Dichloropropane-d6	6.693	67	54678	4.555	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	91.200%
41) Toluene-d8	7.899	98	135081	3.774	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	75.400%
43) trans-1,3-Dichloroprop...	8.182	79	20639	4.090	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.800%
46) 2-Hexanone-d5	8.635	63	120500	46.063	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.120%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	53851	4.492	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	89.800%
66) 1,2-Dichlorobenzene-d4	12.195	152	61353	4.440	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	88.800%
Target Compounds						Qvalue
3) Chloromethane	1.523	50	1268	0.159	ug/L	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031522\
Data File : VU047571.D
Acq On : 15 Mar 2022 12:47
Operator : SY/MD
Sample : N1878-09
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Mar 16 06:45:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR031422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Mar 16 06:43:54 2022
Response via : Initial Calibration





#3
 Chloromethane
 Concen: 0.159 ug/L
 RT: 1.523 min Scan# 51
 Delta R.T. -0.000 min
 Lab File: VU047571.D
 Acq: 15 Mar 2022 12:47

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Tgt Ion: 50 Resp: 1268
 Ion Ratio Lower Upper
 50 100
 52 34.5 23.2 43.0

