

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031524\
 Data File : VU058121.D
 Acq On : 15 Mar 2024 12:47
 Operator : MD/SY
 Sample : P1741-12
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Mar 15 23:30:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR022924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Mar 14 04:37:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	131512	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	125667	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	53291	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	49289	5.209	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	104.200%
7) Chloroethane-d5	1.907	69	42723	5.493	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	109.800%
11) 1,1-Dichloroethene-d2	2.557	65	21651	5.288	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	105.800%
20) 2-Butanone-d5	4.611	46	81054	51.990	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.980%
24) Chloroform-d	5.052	84	93134	5.204	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.000%
26) 1,2-Dichloroethane-d4	5.695	65	44700	5.513	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	110.200%
32) Benzene-d6	5.717	84	189858	5.184	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.600%
36) 1,2-Dichloropropane-d6	6.682	67	59260	5.323	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	106.400%
41) Toluene-d8	7.891	98	158170	4.767	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.400%
43) trans-1,3-Dichloroprop...	8.177	79	16986	4.437	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.800%
46) 2-Hexanone-d5	8.627	63	71304	55.473	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	110.940%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	36455	5.340	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	106.800%
66) 1,2-Dichlorobenzene-d4	12.190	152	49026	5.586	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	111.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031524\
Data File : VU058121.D
Acq On : 15 Mar 2024 12:47
Operator : MD/SY
Sample : P1741-12
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Mar 15 23:30:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR022924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Mar 14 04:37:22 2024
Response via : Initial Calibration

