

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031622\
 Data File : VU047594.D
 Acq On : 16 Mar 2022 16:48
 Operator : SY/MD
 Sample : VIBLK016
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK016

Quant Time: Mar 17 00:52:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR031422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Mar 17 00:50:30 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	124265	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	127033	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	62571	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	26790	3.297	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	66.000%	
7) Chloroethane-d5	1.919	69	30906	4.190	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	83.800%	
11) 1,1-Dichloroethene-d2	2.572	65	10974	2.879	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	57.600%#	
20) 2-Butanone-d5	4.642	46	153708	65.733	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	131.460%#	
24) Chloroform-d	5.067	84	78809	4.416	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	88.400%	
26) 1,2-Dichloroethane-d4	5.703	65	45320	4.687	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	93.800%	
32) Benzene-d6	5.732	84	134933	3.649	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	73.000%	
36) 1,2-Dichloropropane-d6	6.694	67	48606	4.180	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	83.600%	
41) Toluene-d8	7.899	98	104391	3.011	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	60.200%#	
43) trans-1,3-Dichloroprop...	8.182	79	18314	3.746	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	75.000%	
46) 2-Hexanone-d5	8.639	63	129700	51.176	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	102.360%	
56) 1,1,2,2-Tetrachloroeth...	10.758	84	55077	4.743	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	94.800%	
66) 1,2-Dichlorobenzene-d4	12.195	152	54969	4.204	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	84.000%	
Target Compounds						
					Qvalue	
5) Vinyl chloride	1.607	62	622	0.080	ug/L #	26
22) cis-1,2-Dichloroethene	4.674	96	695	0.094	ug/L	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031622\
Data File : VU047594.D
Acq On : 16 Mar 2022 16:48
Operator : SY/MD
Sample : VIBLK016
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK016

Quant Time: Mar 17 00:52:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR031422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Mar 17 00:50:30 2022
Response via : Initial Calibration

