

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031824\
 Data File : VU058149.D
 Acq On : 18 Mar 2024 15:11
 Operator : MD/SY
 Sample : P1779-06
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Mar 19 01:38:16 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR022924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Mar 14 04:37:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	134634	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	131209	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	54927	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	42496	4.387	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	87.800%
7) Chloroethane-d5	1.904	69	39248	4.930	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	98.600%
11) 1,1-Dichloroethene-d2	2.557	65	19085	4.553	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	91.000%
20) 2-Butanone-d5	4.615	46	78923	49.449	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	98.900%
24) Chloroform-d	5.052	84	87509	4.776	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.600%
26) 1,2-Dichloroethane-d4	5.695	65	40624	4.894	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.800%
32) Benzene-d6	5.718	84	176098	4.605	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.200%
36) 1,2-Dichloropropane-d6	6.682	67	55273	4.755	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.200%
41) Toluene-d8	7.894	98	149670	4.320	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.400%
43) trans-1,3-Dichloroprop...	8.177	79	15646	3.914	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	78.200%
46) 2-Hexanone-d5	8.627	63	63651	47.427	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	94.860%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	33076	4.640	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.800%
66) 1,2-Dichlorobenzene-d4	12.190	152	44368	4.905	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031824\
Data File : VU058149.D
Acq On : 18 Mar 2024 15:11
Operator : MD/SY
Sample : P1779-06
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Mar 19 01:38:16 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR022924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Mar 14 04:37:22 2024
Response via : Initial Calibration

