

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031925\
 Data File : VU063306.D
 Acq On : 19 Mar 2025 12:25
 Operator : MD/SY
 Sample : VSTD0.551
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VSTD0.5051

Quant Time: Mar 25 09:55:28 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM031925.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Mar 25 09:54:25 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.258	114	11478	0.50	ug/L	0.00
5) Chlorobenzene-d5	9.419	117	15355	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.808	152	4847	0.50	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.590	65	3748	0.51	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.685	65	2783	0.45	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.687	67	2485	0.50	ug/L	0.00
8) Toluene-d8	7.887	98	6065	0.45	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.745	84	1871	0.44	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.597	62	6955	0.52	ug/L	100
6) Trichloroethene	6.544	95	5196	0.45	ug/L	100
9) 1,2-Dibromoethane	8.924	107	2970	0.48	ug/L	100
12) 1,2,3-Trichloropropane	10.814	75	2613	0.46	ug/L	100
13) 1,2-Dibromo-3-chloropr...	12.990	75	586	0.44	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031925\
Data File : VU063306.D
Acq On : 19 Mar 2025 12:25
Operator : MD/SY
Sample : VSTD0.551
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTD0.5051

Quant Time: Mar 25 09:55:28 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM031925.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Mar 25 09:54:25 2025
Response via : Initial Calibration

