

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032822\
 Data File : VU047730.D
 Acq On : 28 Mar 2022 16:15
 Operator : SY/MD
 Sample : N2081-18DL 20X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EW5Y5DL

Quant Time: Mar 29 01:45:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Mar 29 01:43:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	313991	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	322207	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.813	152	177788	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	85845	37.702	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	75.400%	
7) Chloroethane-d5	1.919	69	78400	43.668	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	87.340%	
11) 1,1-Dichloroethene-d2	2.572	63	128422	30.827	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	61.660%	
21) 2-Butanone-d5	4.642	46	228645	116.992	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	116.990%	
24) Chloroform-d	5.067	84	192078	46.104	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	92.200%	
26) 1,2-Dichloroethane-d4	5.703	65	127165	49.767	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	99.540%	
32) Benzene-d6	5.729	84	381950	47.110	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	94.220%	
36) 1,2-Dichloropropane-d6	6.694	67	123722	49.689	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	99.380%	
41) Toluene-d8	7.899	98	345189	43.699	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	87.400%	
43) trans-1,3-Dichloroprop...	8.179	79	56896	45.104	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	90.200%	
47) 2-Hexanone-d5	8.636	63	164066	112.559	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	112.560%	
56) 1,1,2,2-Tetrachloroeth...	10.758	84	212628	49.759	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	99.520%	
66) 1,2-Dichlorobenzene-d4	12.192	152	157855	49.906	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	99.820%	
Target Compounds						
					Qvalue	
5) Vinyl chloride	1.607	62	4056	1.429	ug/L #	53
8) Chloroethane	1.942	64	6457	3.623	ug/L	93
12) 1,1-Dichloroethene	2.588	96	3711	1.723	ug/L #	1
19) 1,1-Dichloroethane	3.874	63	91559	22.852	ug/L	99
20) cis-1,2-Dichloroethene	4.671	96	185271	72.744	ug/L	98
30) 1,1,1-Trichloroethane	5.321	97	91023	24.914	ug/L	98
46) Tetrachloroethene	8.552	164	2625	1.195	ug/L	83
54) o-Xylene	10.102	106	4054	0.978	ug/L	95
62) 1,3,5-Trimethylbenzene	11.089	105	6838	1.312	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032822\
Data File : VU047730.D
Acq On : 28 Mar 2022 16:15
Operator : SY/MD
Sample : N2081-18DL 20X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EW5Y5DL

Quant Time: Mar 29 01:45:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Mar 29 01:43:13 2022
Response via : Initial Calibration

