

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032822\
 Data File : VU047741.D
 Acq On : 28 Mar 2022 20:23
 Operator : SY/MD
 Sample : N2120-02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0HS9

Quant Time: Mar 29 01:47:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Mar 29 01:43:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	343660	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	353269	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	171918	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	91285	36.630	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	73.260%
7) Chloroethane-d5	1.919	69	82241	41.852	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	83.700%
11) 1,1-Dichloroethene-d2	2.572	63	127521	27.968	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	55.940%#
21) 2-Butanone-d5	4.642	46	249534	116.658	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	116.660%
24) Chloroform-d	5.067	84	205809	45.135	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.260%
26) 1,2-Dichloroethane-d4	5.703	65	135031	48.283	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.560%
32) Benzene-d6	5.729	84	398615	44.842	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.680%
36) 1,2-Dichloropropane-d6	6.690	67	130447	47.783	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.560%
41) Toluene-d8	7.899	98	354280	40.906	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	81.820%
43) trans-1,3-Dichloroprop...	8.179	79	60222	43.543	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.080%
47) 2-Hexanone-d5	8.636	63	173130	108.333	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	108.330%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	208224	44.444	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	88.880%
66) 1,2-Dichlorobenzene-d4	12.192	152	150008	49.045	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.080%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032822\
Data File : VU047741.D
Acq On : 28 Mar 2022 20:23
Operator : SY/MD
Sample : N2120-02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0HS9

Quant Time: Mar 29 01:47:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Mar 29 01:43:13 2022
Response via : Initial Calibration

