

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032822\
 Data File : VU047737.D
 Acq On : 28 Mar 2022 18:53
 Operator : SY/MD
 Sample : VIBLK054
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK054

Quant Time: Mar 29 01:47:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Mar 29 01:43:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	340699	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	323266	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	187542	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	96117	38.904	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	77.800%
7) Chloroethane-d5	1.919	69	86164	44.230	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	88.460%
11) 1,1-Dichloroethene-d2	2.572	63	131862	29.172	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	58.340%#
21) 2-Butanone-d5	4.639	46	249606	117.706	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	117.710%
24) Chloroform-d	5.067	84	205386	45.433	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.860%
26) 1,2-Dichloroethane-d4	5.703	65	135811	48.984	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.960%
32) Benzene-d6	5.729	84	410637	50.482	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.960%
36) 1,2-Dichloropropane-d6	6.690	67	123135	49.291	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	98.580%
41) Toluene-d8	7.899	98	335137	42.288	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	84.580%
43) trans-1,3-Dichloroprop...	8.179	79	56619	44.737	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.480%
47) 2-Hexanone-d5	8.636	63	162694	111.252	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	111.250%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	229273	53.479	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	106.960%
66) 1,2-Dichlorobenzene-d4	12.195	152	168654	50.547	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.100%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032822\
Data File : VU047737.D
Acq On : 28 Mar 2022 18:53
Operator : SY/MD
Sample : VIBLK054
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK054

Quant Time: Mar 29 01:47:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Mar 29 01:43:13 2022
Response via : Initial Calibration

