

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032922\
 Data File : VU047753.D
 Acq On : 29 Mar 2022 16:49
 Operator : SY/MD
 Sample : N2111-01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0PC9

Quant Time: Mar 30 01:23:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Mar 30 01:20:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	310071	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	318315	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	171162	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	71979	32.012	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	64.020%
7) Chloroethane-d5	1.919	69	69005	38.921	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	77.840%
11) 1,1-Dichloroethene-d2	2.571	63	107824	26.210	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	52.420%#
21) 2-Butanone-d5	4.642	46	237258	122.934	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	122.930%
24) Chloroform-d	5.066	84	182049	44.249	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.500%
26) 1,2-Dichloroethane-d4	5.703	65	122091	48.385	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.760%
32) Benzene-d6	5.729	84	342489	42.759	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.520%
36) 1,2-Dichloropropane-d6	6.693	67	117503	47.768	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.540%
41) Toluene-d8	7.899	98	308331	39.510	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	79.020%#
43) trans-1,3-Dichloroprop...	8.179	79	53172	42.667	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.340%
47) 2-Hexanone-d5	8.635	63	167034	115.996	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	116.000%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	212539	50.346	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	100.700%
66) 1,2-Dichlorobenzene-d4	12.192	152	145712	47.851	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.700%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032922\
Data File : VU047753.D
Acq On : 29 Mar 2022 16:49
Operator : SY/MD
Sample : N2111-01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0PC9

Quant Time: Mar 30 01:23:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Mar 30 01:20:49 2022
Response via : Initial Calibration

