

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU033020\
 Data File : VU037466.D
 Acq On : 30 Mar 2020 16:36
 Operator : JC/MD
 Sample : VIBLK28
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK28

Quant Time: Mar 31 02:52:30 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR032820WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Mar 31 01:16:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	155044	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	145426	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	74506	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	31920	2.79	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	55.80%
7) Chloroethane-d5	1.93	69	28930	3.37	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	67.40%
11) 1,1-Dichloroethene-d2	2.59	63	50363	2.25	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	45.00%#
20) 2-Butanone-d5	4.67	46	108196	34.68	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	69.36%
24) Chloroform-d	5.10	84	67121	2.97	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	59.40%#
26) 1,2-Dichloroethane-d4	5.74	65	39912	3.21	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	64.20%#
32) Benzene-d6	5.76	84	133403	3.18	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	63.60%#
36) 1,2-Dichloropropane-d6	6.72	67	44825	3.22	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	64.40%
41) Toluene-d8	7.93	98	122080	3.06	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	61.20%#
43) trans-1,3-Dichloropropene-	8.20	79	19876	3.04	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	60.80%
46) 2-Hexanone-d5	8.66	63	89329	34.36	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	68.72%
57) 1,1,2,2-Tetrachloroethane-	10.78	84	36199	3.10	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	62.00%#
64) 1,2-Dichlorobenzene-d4	12.21	152	45706	3.11	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	62.20%#

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU033020\
Data File : VU037466.D
Acq On : 30 Mar 2020 16:36
Operator : JC/MD
Sample : VIBLK28
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK28

Quant Time: Mar 31 02:52:30 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR032820WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Mar 31 01:16:50 2020
Response via : Initial Calibration

