

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU033020\
 Data File : VU037489.D
 Acq On : 31 Mar 2020 04:01
 Operator : JC/MD
 Sample : L2064-06
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0Y02

Quant Time: Mar 31 06:45:37 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR032820WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Mar 31 04:01:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	98460	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	93956	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	45784	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	28284	3.89	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	77.80%
7) Chloroethane-d5	1.93	69	26648	4.89	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	97.80%
11) 1,1-Dichloroethene-d2	2.59	63	42913	3.02	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	60.40%
20) 2-Butanone-d5	4.69	46	106231	53.62	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	107.24%
24) Chloroform-d	5.10	84	60147	4.19	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.80%
26) 1,2-Dichloroethane-d4	5.74	65	34764	4.40	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.00%
32) Benzene-d6	5.76	84	120229	4.43	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.60%
36) 1,2-Dichloropropane-d6	6.72	67	40302	4.49	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	89.80%
41) Toluene-d8	7.92	98	109677	4.26	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.20%
43) trans-1,3-Dichloropropene-	8.20	79	16981	4.02	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.40%
46) 2-Hexanone-d5	8.66	63	90279	53.75	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	107.50%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	34742	4.61	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.20%
64) 1,2-Dichlorobenzene-d4	12.21	152	39688	4.40	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	88.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU033020\
Data File : VU037489.D
Acq On : 31 Mar 2020 04:01
Operator : JC/MD
Sample : L2064-06
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0Y02

Quant Time: Mar 31 06:45:37 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR032820WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Mar 31 04:01:00 2020
Response via : Initial Calibration

