

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU033020\
 Data File : VU037485.D
 Acq On : 31 Mar 2020 02:23
 Operator : JC/MD
 Sample : L2063-02
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFCJ5

Quant Time: Mar 31 06:27:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR032820WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Mar 31 04:01:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	101264	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	93920	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	45209	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	28817	3.86	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	77.20%
7) Chloroethane-d5	1.93	69	26272	4.69	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.80%
11) 1,1-Dichloroethene-d2	2.59	63	45059	3.09	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	61.80%
20) 2-Butanone-d5	4.69	46	112055	55.00	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	110.00%
24) Chloroform-d	5.11	84	61141	4.14	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.80%
26) 1,2-Dichloroethane-d4	5.74	65	35151	4.33	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.60%
32) Benzene-d6	5.76	84	123024	4.53	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.60%
36) 1,2-Dichloropropane-d6	6.72	67	42032	4.68	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.60%
41) Toluene-d8	7.92	98	110379	4.29	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.80%
43) trans-1,3-Dichloropropene-	8.21	79	17671	4.18	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.60%
46) 2-Hexanone-d5	8.66	63	91753	54.65	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	109.30%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	35718	4.74	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.80%
64) 1,2-Dichlorobenzene-d4	12.21	152	41506	4.66	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU033020\
Data File : VU037485.D
Acq On : 31 Mar 2020 02:23
Operator : JC/MD
Sample : L2063-02
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFCJ5

Quant Time: Mar 31 06:27:17 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR032820WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Mar 31 04:01:00 2020
Response via : Initial Calibration

