

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040224\
 Data File : VU058343.D
 Acq On : 02 Apr 2024 17:44
 Operator : MD/SY
 Sample : P1912-08
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DCOD9

Quant Time: Apr 03 02:04:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR032124WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Apr 03 01:56:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	75936	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	72790	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	33521	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	19827	3.090	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	61.800%
7) Chloroethane-d5	1.908	69	20576	3.928	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	78.600%
11) 1,1-Dichloroethene-d2	2.560	65	8832	3.315	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	66.200%
20) 2-Butanone-d5	4.621	46	50363	44.177	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	88.360%
24) Chloroform-d	5.052	84	44518	3.880	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	77.600%
26) 1,2-Dichloroethane-d4	5.695	65	22401	4.176	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.600%
32) Benzene-d6	5.718	84	91419	3.867	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	77.400%
36) 1,2-Dichloropropane-d6	6.682	67	29266	4.200	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	84.000%
41) Toluene-d8	7.891	98	75018	3.582	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	71.600%
43) trans-1,3-Dichloroprop...	8.174	79	8317	3.725	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	74.600%
46) 2-Hexanone-d5	8.627	63	37871	41.885	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	83.780%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	19783	4.358	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	87.200%
66) 1,2-Dichlorobenzene-d4	12.187	152	26746	4.285	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	85.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040224\
Data File : VU058343.D
Acq On : 02 Apr 2024 17:44
Operator : MD/SY
Sample : P1912-08
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DCOD9

Quant Time: Apr 03 02:04:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR032124WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Apr 03 01:56:26 2024
Response via : Initial Calibration

