

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040224\
 Data File : VU058326.D
 Acq On : 02 Apr 2024 09:53
 Operator : MD/SY
 Sample : VU0402WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK160

Quant Time: Apr 03 02:01:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR032124WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Apr 03 01:56:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	71680	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	68817	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	28172	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	27793	4.588	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	91.800%
7) Chloroethane-d5	1.907	69	24970	5.050	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	101.000%
11) 1,1-Dichloroethene-d2	2.560	65	12242	4.868	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	97.400%
20) 2-Butanone-d5	4.628	46	55054	51.159	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	102.320%
24) Chloroform-d	5.055	84	52194	4.820	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.400%
26) 1,2-Dichloroethane-d4	5.692	65	25339	5.005	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.000%
32) Benzene-d6	5.721	84	105701	4.729	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.600%
36) 1,2-Dichloropropane-d6	6.682	67	32978	5.006	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.200%
41) Toluene-d8	7.891	98	87787	4.433	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.600%
43) trans-1,3-Dichloroprop...	8.174	79	9383	4.445	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.000%
46) 2-Hexanone-d5	8.631	63	39491	46.198	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.400%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	20731	4.831	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.600%
66) 1,2-Dichlorobenzene-d4	12.187	152	26163	4.987	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040224\
Data File : VU058326.D
Acq On : 02 Apr 2024 09:53
Operator : MD/SY
Sample : VU0402WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK160

Quant Time: Apr 03 02:01:03 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR032124WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Apr 03 01:56:26 2024
Response via : Initial Calibration

