

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040224\
 Data File : VU058328.D
 Acq On : 02 Apr 2024 10:44
 Operator : MD/SY
 Sample : P1912-14
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DCOE4

Quant Time: Apr 03 02:01:29 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR032124WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Apr 03 01:56:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	62994	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	59982	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	26754	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	23730	4.458	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	89.200%	
7) Chloroethane-d5	1.907	69	22908	5.272	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	105.400%	
11) 1,1-Dichloroethene-d2	2.563	65	10369	4.691	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	93.800%	
20) 2-Butanone-d5	4.628	46	53875	56.966	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	113.940%	
24) Chloroform-d	5.052	84	49386	5.189	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	103.800%	
26) 1,2-Dichloroethane-d4	5.695	65	24711	5.554	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	111.000%	
32) Benzene-d6	5.718	84	99764	5.121	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	102.400%	
36) 1,2-Dichloropropane-d6	6.682	67	31483	5.483	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	109.600%	
41) Toluene-d8	7.891	98	83238	4.823	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	96.400%	
43) trans-1,3-Dichloroprop...	8.174	79	9077	4.934	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	98.600%	
46) 2-Hexanone-d5	8.631	63	38423	51.570	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	103.140%	
56) 1,1,2,2-Tetrachloroeth...	10.750	84	20640	5.518	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	110.400%	
66) 1,2-Dichlorobenzene-d4	12.187	152	26853	5.390	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	107.800%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040224\
Data File : VU058328.D
Acq On : 02 Apr 2024 10:44
Operator : MD/SY
Sample : P1912-14
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DCOE4

Quant Time: Apr 03 02:01:29 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR032124WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Apr 03 01:56:26 2024
Response via : Initial Calibration

