

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040524\
 Data File : VU058421.D
 Acq On : 05 Apr 2024 16:00
 Operator : MD/SY
 Sample : P1925-13
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PMW-1(20240327)

Quant Time: Apr 05 23:36:57 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR040424WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Apr 05 23:33:02 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	75882	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	74876	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	33005	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	23553	4.688	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	93.800%
7) Chloroethane-d5	1.907	69	21454	4.914	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	98.200%
11) 1,1-Dichloroethene-d2	2.560	65	10471	4.605	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	92.000%
20) 2-Butanone-d5	4.627	46	52935	49.999	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	100.000%
24) Chloroform-d	5.055	84	48632	4.849	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.000%
26) 1,2-Dichloroethane-d4	5.695	65	24114	5.152	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.000%
32) Benzene-d6	5.721	84	97226	4.823	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.400%
36) 1,2-Dichloropropane-d6	6.682	67	31181	4.932	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.600%
41) Toluene-d8	7.891	98	81389	4.553	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.000%
43) trans-1,3-Dichloroprop...	8.174	79	8915	4.543	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.800%
46) 2-Hexanone-d5	8.630	63	37283	46.063	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.120%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	19598	4.391	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	87.800%
66) 1,2-Dichlorobenzene-d4	12.190	152	25387	4.567	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	91.400%
Target Compounds						
					Qvalue	
16) Methylene chloride	3.039	84	3351	0.470	ug/L	97
22) cis-1,2-Dichloroethene	4.656	96	791	0.125	ug/L	94
70) 1,2,4-trichlorobenzene	13.842	180	1533	0.229	ug/L	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040524\
Data File : VU058421.D
Acq On : 05 Apr 2024 16:00
Operator : MD/SY
Sample : P1925-13
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PMW-1(20240327)

Quant Time: Apr 05 23:36:57 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR040424WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Apr 05 23:33:02 2024
Response via : Initial Calibration









#16
Methylene chloride
Concen: 0.470 ug/L
RT: 3.039 min Scan# 544
Delta R.T. -0.000 min
Lab File: VU058421.D
Acq: 05 Apr 2024 16:00




Instrument :
MSVOA_U
ClientSampleId :
PMW-1(20240327)


Tgt Ion: 84 Resp: 3351
Ion Ratio Lower Upper
84 100
86 68.9 44.3 82.3
49 135.2 96.0 178.2



#22
cis-1,2-Dichloroethene
Concen: 0.125 ug/L
RT: 4.656 min Scan# 1047
Delta R.T. -0.003 min
Lab File: VU058421.D
Acq: 05 Apr 2024 16:00

Tgt Ion: 96 Resp: 791
Ion Ratio Lower Upper
96 100
61 143.5 98.1 182.3
98 51.4 43.6 81.0





#70

1,2,4-trichlorobenzene

Concen: 0.229 ug/L

RT: 13.842 min Scan# 3904

Delta R.T. 0.009 min

Lab File: VU058421.D


Acq: 05 Apr 2024 16:00

Instrument :

MSVOA_U

ClientSampleId :

PMW-1(20240327)



Tgt Ion:180 Resp: 1533

Ion Ratio Lower Upper

180 100

182 91.0 74.4 111.6

145 26.4 24.1 36.1

