

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040524\
 Data File : VU058441.D
 Acq On : 06 Apr 2024 00:11
 Operator : MD/SY
 Sample : P1948-10
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 COAL3

Quant Time: Apr 06 05:59:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR040424WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Apr 06 00:37:11 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	80947	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	79209	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	34452	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	22165	4.136	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	82.800%
7) Chloroethane-d5	1.907	69	20103	4.316	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	86.400%
11) 1,1-Dichloroethene-d2	2.563	65	10527	4.340	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	86.800%
20) 2-Butanone-d5	4.628	46	56626	50.138	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	100.280%
24) Chloroform-d	5.055	84	48633	4.546	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.000%
26) 1,2-Dichloroethane-d4	5.695	65	24216	4.850	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.000%
32) Benzene-d6	5.718	84	97588	4.576	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.600%
36) 1,2-Dichloropropane-d6	6.682	67	31601	4.725	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.600%
41) Toluene-d8	7.891	98	80935	4.280	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.600%
43) trans-1,3-Dichloroprop...	8.177	79	8689	4.186	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.800%
46) 2-Hexanone-d5	8.627	63	38880	45.408	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	90.820%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	19946	4.225	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	84.400%
66) 1,2-Dichlorobenzene-d4	12.187	152	25315	4.363	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	87.200%
Target Compounds					Qvalue	
13) Acetone	2.650	43	6030	6.256	ug/L	93


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040524\
Data File : VU058441.D
Acq On : 06 Apr 2024 00:11
Operator : MD/SY
Sample : P1948-10
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
COAL3

Quant Time: Apr 06 05:59:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR040424WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Apr 06 00:37:11 2024
Response via : Initial Calibration





#13
 Acetone
 Concen: 6.256 ug/L
 RT: 2.650 min Scan# 41
 Delta R.T. 0.013 min
 Lab File: VU058441.D
 Acq: 06 Apr 2024 00:11

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AL3

Tgt Ion: 43 Resp: 6030
 Ion Ratio Lower Upper
 43 100
 58 26.7 0.0 61.4

