

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040824\
 Data File : VU058457.D
 Acq On : 08 Apr 2024 10:29
 Operator : MD/SY
 Sample : VU0408WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK165

Quant Time: Apr 09 04:22:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR040424WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Apr 09 04:22:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	77571	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	75956	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	31230	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	18038	3.512	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	70.200%
7) Chloroethane-d5	1.907	69	17892	4.009	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	80.200%
11) 1,1-Dichloroethene-d2	2.560	65	8469	3.643	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.800%
20) 2-Butanone-d5	4.628	46	55916	51.664	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.320%
24) Chloroform-d	5.052	84	47473	4.631	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.600%
26) 1,2-Dichloroethane-d4	5.695	65	23283	4.866	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.400%
32) Benzene-d6	5.718	84	88047	4.306	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.200%
36) 1,2-Dichloropropane-d6	6.682	67	30131	4.698	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.000%
41) Toluene-d8	7.891	98	73247	4.039	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.800%
43) trans-1,3-Dichloroprop...	8.174	79	8610	4.325	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.600%
46) 2-Hexanone-d5	8.631	63	36142	44.018	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	88.040%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	20393	4.504	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.000%
66) 1,2-Dichlorobenzene-d4	12.187	152	24804	4.716	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	94.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040824\
Data File : VU058457.D
Acq On : 08 Apr 2024 10:29
Operator : MD/SY
Sample : VU0408WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK165

Quant Time: Apr 09 04:22:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR040424WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Apr 09 04:22:28 2024
Response via : Initial Calibration

