



Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040824\
Data File : VU058455.D
Acq On : 08 Apr 2024 08:25
Operator : MD/SY
Sample : BFB138
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFB138

Integration File: rteint.p

Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR040424WMA.M
Title : TRACE VOA SFAM1.0
Last Update : Wed Apr 10 02:08:42 2024



AutoFind: Scans 2903, 2904, 2905; Background Corrected with Scan 2890

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.9	18451	PASS
75	95	30	80	56.4	45413	PASS
95	95	100	100	100.0	80549	PASS
96	95	5	9	6.7	5422	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	76.4	61523	PASS
175	174	5	9	7.8	4780	PASS
176	174	95	101	97.2	59829	PASS
177	176	5	9	6.2	3700	PASS

m/z	Abundance
35.90	447.0
37.90	179.0
39.80	102.0
43.70	106.0

Instrument :
MSVOA_U
ClientSampleId :
BFB138

m/z	Abundance
35.90	1414.0
36.90	5083.0
37.90	4490.0
38.90	1856.0
39.80	157.0
43.90	810.0
45.00	750.0
46.00	111.0
46.90	1093.0
47.90	619.0
48.90	3962.0
50.00	18128.0
51.00	5679.0
51.80	265.0
54.90	274.0
55.90	1491.0
56.90	2640.0
59.90	835.0
60.90	4796.0
61.90	4446.0
62.90	3519.0
63.90	479.0
64.90	621.0
66.90	248.0
67.90	9204.0
68.90	9784.0
69.90	636.0
71.90	453.0
72.90	4182.0
73.90	15430.0
75.00	43912.0
75.90	3841.0
76.90	523.0
77.90	356.0
78.80	3529.0
79.90	919.0
80.80	3290.0
81.80	768.0
85.80	102.0
86.90	3554.0
87.80	3433.0
90.80	406.0
91.90	2634.0
92.90	3965.0
93.90	9798.0
95.00	77944.0
95.90	5345.0
96.90	133.0
103.80	452.0
104.80	135.0
105.80	460.0
115.80	288.0
116.80	680.0
117.80	370.0
118.80	566.0
127.80	312.0
128.80	104.0
129.90	271.0
130.70	139.0
134.70	155.0
136.80	156.0
140.90	898.0
141.80	123.0
142.80	929.0
144.80	159.0
145.80	105.0
147.80	179.0
154.90	217.0
156.80	112.0
171.80	657.0
173.90	57656.0
174.90	4475.0
175.90	55960.0
176.90	3447.0

Instrument :
MSVOA_U
ClientSampleId :
BFB138

m/z	Abundance
35.90	1487.0
36.90	5399.0
37.90	4696.0
38.90	1912.0
39.80	160.0
43.90	845.0
44.90	829.0
45.90	105.0
46.90	1243.0
47.80	677.0
48.90	4186.0
50.00	19040.0
51.00	6118.0
51.80	258.0
54.90	309.0
55.90	1623.0
56.90	2835.0
57.90	110.0
59.90	912.0
60.90	5068.0
61.90	4641.0
63.00	3738.0
63.90	523.0
65.00	691.0
66.90	303.0
67.90	9909.0
68.90	10400.0
69.90	681.0
71.90	492.0
72.90	4467.0
73.90	16269.0
75.00	46832.0
75.90	4091.0
76.90	545.0
77.90	393.0
78.80	3777.0
79.90	1005.0
80.80	3644.0
81.80	776.0
85.70	112.0
86.90	3924.0
87.80	3536.0
90.80	478.0
91.90	2710.0
92.90	4327.0
93.90	10613.0
94.90	83056.0
96.00	5639.0
96.90	135.0
103.80	443.0
104.80	151.0
105.80	474.0
115.80	328.0
116.80	717.0
117.80	421.0
118.80	597.0
127.80	320.0
128.80	114.0
129.90	305.0
130.70	167.0
134.70	153.0
136.80	163.0
140.80	966.0
141.70	123.0
142.90	1009.0
144.70	148.0
145.90	110.0
147.80	234.0
154.90	244.0
156.80	139.0
158.70	114.0
171.80	724.0
173.90	63456.0
174.90	4934.0
175.90	61824.0
176.90	3854.0

Instrument :
MSVOA_U
ClientSampleId :
BFB138

m/z	Abundance
35.90	1413.0
36.90	5086.0
37.90	4385.0
38.90	1814.0
39.80	158.0
43.90	797.0
44.90	810.0
46.90	1179.0
47.80	656.0
48.90	4015.0
50.00	18184.0
51.00	5948.0
51.90	217.0
54.90	321.0
55.90	1552.0
56.90	2771.0
57.90	126.0
59.90	883.0
61.00	4884.0
61.90	4461.0
63.00	3707.0
63.90	524.0
65.00	670.0
66.90	318.0
67.90	9562.0
68.90	9854.0
69.90	670.0
72.00	459.0
72.90	4395.0
73.90	15522.0
75.00	45496.0
75.90	3929.0
76.90	536.0
77.90	382.0
78.80	3646.0
79.90	978.0
80.80	3563.0
81.80	728.0
86.00	102.0
86.90	3752.0
87.80	3327.0
90.80	452.0
91.90	2463.0
92.90	4155.0
93.90	10388.0
94.90	80648.0
96.00	5282.0
96.90	130.0
103.70	424.0
104.80	142.0
105.80	431.0
115.80	335.0
116.80	704.0
117.80	426.0
118.80	558.0
127.80	309.0
128.80	141.0
129.80	316.0
130.70	178.0
134.70	164.0
136.80	143.0
140.80	921.0
141.70	109.0
142.90	1028.0
144.80	129.0
145.90	117.0
147.80	266.0
154.90	230.0
156.80	139.0
158.80	112.0
171.80	740.0
173.90	63456.0
174.90	4930.0
175.90	61704.0
176.90	3800.0

Instrument :
MSVOA_U
ClientSampleId :
BFB138