

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU041520\
 Data File : VU037726.D
 Acq On : 15 Apr 2020 12:11
 Operator : JC/MD
 Sample : L2275-17DL 8X
 Misc : 25mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 YAQD7DL

Quant Time: Apr 16 11:33:16 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR041420WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Apr 16 01:23:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	252772	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	251622	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	131591	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	92844	4.85	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	97.00%
7) Chloroethane-d5	1.90	69	73944	4.58	ug/L	-0.03
Spiked Amount	5.000	Range	65 - 130	Recovery	=	91.60%
11) 1,1-Dichloroethene-d2	2.58	63	116874	3.44	ug/L	-0.01
Spiked Amount	5.000	Range	60 - 125	Recovery	=	68.80%
20) 2-Butanone-d5	4.66	46	275693	52.48	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.96%
24) Chloroform-d	5.10	84	141713	4.64	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.80%
26) 1,2-Dichloroethane-d4	5.74	65	87940	5.01	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.20%
32) Benzene-d6	5.76	84	319819	4.69	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.80%
36) 1,2-Dichloropropane-d6	6.72	67	105529	4.77	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.40%
41) Toluene-d8	7.92	98	292529	4.74	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.80%
43) trans-1,3-Dichloropropene-	8.20	79	46160	4.64	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.80%
46) 2-Hexanone-d5	8.66	63	282925	55.61	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	111.22%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	92752	5.31	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	106.20%
64) 1,2-Dichlorobenzene-d4	12.21	152	118176	5.03	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.60%

Target Compounds					Ovalue
13) Acetone	2.66	43	3261	0.942 ug/L	98
47) Tetrachloroethene	8.57	164	103817	7.071 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU041520\
Data File : VU037726.D
Acq On : 15 Apr 2020 12:11
Operator : JC/MD
Sample : L2275-17DL 8X
Misc : 25mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
YAQD7DL

Quant Time: Apr 16 11:33:16 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR041420WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Apr 16 01:23:00 2020
Response via : Initial Calibration

