

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041624\
 Data File : VU058709.D
 Acq On : 16 Apr 2024 12:34
 Operator : MD/SY
 Sample : P2055-03DL 20X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0BL3DL

Quant Time: Apr 17 01:11:23 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR040424WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Apr 17 01:09:43 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	69087	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	67627	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	28982	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	15455	3.379	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	67.600%	
7) Chloroethane-d5	1.904	69	18049	4.541	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	90.800%	
11) 1,1-Dichloroethene-d2	2.560	65	6992	3.377	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	67.600%	
20) 2-Butanone-d5	4.615	46	62936	65.292	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	130.580%#	
24) Chloroform-d	5.052	84	45289	4.960	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	99.200%	
26) 1,2-Dichloroethane-d4	5.692	65	24909	5.845	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	117.000%	
32) Benzene-d6	5.718	84	78825	4.329	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	86.600%	
36) 1,2-Dichloropropane-d6	6.682	67	29789	5.217	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	104.400%	
41) Toluene-d8	7.891	98	57250	3.546	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	71.000%	
43) trans-1,3-Dichloroprop...	8.177	79	7502	4.233	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	84.600%	
46) 2-Hexanone-d5	8.627	63	41236	56.408	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	112.820%	
56) 1,1,2,2-Tetrachloroeth...	10.749	84	21159	5.249	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	105.000%	
66) 1,2-Dichlorobenzene-d4	12.190	152	21117	4.326	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	86.600%	
Target Compounds						
					Qvalue	
16) Methylene chloride	3.036	84	7029	1.082	ug/L	98
47) Tetrachloroethene	8.547	164	48341	10.671	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041624\
Data File : VU058709.D
Acq On : 16 Apr 2024 12:34
Operator : MD/SY
Sample : P2055-03DL 20X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0BL3DL

Quant Time: Apr 17 01:11:23 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR040424WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Apr 17 01:09:43 2024
Response via : Initial Calibration

