

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041624\
 Data File : VU058717.D
 Acq On : 16 Apr 2024 15:39
 Operator : MD/SY
 Sample : P2058-04
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0BQ8

Quant Time: Apr 17 01:13:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR040424WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Apr 17 01:09:43 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	73202	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	72927	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	30617	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	15899	3.281	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	65.600%
7) Chloroethane-d5	1.904	69	19804	4.702	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.000%
11) 1,1-Dichloroethene-d2	2.557	65	7423	3.384	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.600%
20) 2-Butanone-d5	4.618	46	67300	65.894	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	131.780%#
24) Chloroform-d	5.052	84	47483	4.908	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.200%
26) 1,2-Dichloroethane-d4	5.692	65	24839	5.501	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	110.000%
32) Benzene-d6	5.718	84	81527	4.152	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.000%
36) 1,2-Dichloropropane-d6	6.682	67	30308	4.922	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.400%
41) Toluene-d8	7.891	98	59767	3.433	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	68.600%#
43) trans-1,3-Dichloroprop...	8.177	79	7792	4.077	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.600%
46) 2-Hexanone-d5	8.627	63	43710	55.447	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	110.900%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	21508	4.948	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.000%
66) 1,2-Dichlorobenzene-d4	12.187	152	22142	4.294	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	85.800%
Target Compounds						
					Qvalue	
47) Tetrachloroethene	8.550	164	2006	0.411	ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041624\
Data File : VU058717.D
Acq On : 16 Apr 2024 15:39
Operator : MD/SY
Sample : P2058-04
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0BQ8

Quant Time: Apr 17 01:13:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR040424WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Apr 17 01:09:43 2024
Response via : Initial Calibration

