

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041624\
 Data File : VU058715.D
 Acq On : 16 Apr 2024 14:53
 Operator : MD/SY
 Sample : P2058-02
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0BQ5

Quant Time: Apr 17 01:12:42 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR040424WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Apr 17 01:09:43 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	83719	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	82116	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	33863	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	28765	5.190	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	103.800%
7) Chloroethane-d5	1.904	69	23841	4.949	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.000%
11) 1,1-Dichloroethene-d2	2.557	65	11755	4.686	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	93.800%
20) 2-Butanone-d5	4.615	46	57290	49.047	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	98.100%
24) Chloroform-d	5.052	84	52984	4.789	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.800%
26) 1,2-Dichloroethane-d4	5.692	65	26887	5.207	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.200%
32) Benzene-d6	5.718	84	102567	4.639	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.800%
36) 1,2-Dichloropropane-d6	6.682	67	33201	4.789	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.800%
41) Toluene-d8	7.891	98	83441	4.256	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.200%
43) trans-1,3-Dichloroprop...	8.174	79	10424	4.844	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	96.800%
46) 2-Hexanone-d5	8.627	63	34942	39.364	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	78.720%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	21073	4.305	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.200%
66) 1,2-Dichlorobenzene-d4	12.187	152	25172	4.414	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	88.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041624\
Data File : VU058715.D
Acq On : 16 Apr 2024 14:53
Operator : MD/SY
Sample : P2058-02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0BQ5

Quant Time: Apr 17 01:12:42 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR040424WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Apr 17 01:09:43 2024
Response via : Initial Calibration

