

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041824\
 Data File : VU058759.D
 Acq On : 18 Apr 2024 13:51
 Operator : MD/SY
 Sample : VIBLK008
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK008

Quant Time: Apr 19 06:55:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR040424WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Apr 18 05:04:53 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	66954	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	63438	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	28674	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	17573	3.964	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	79.200%
7) Chloroethane-d5	1.907	69	14517	3.768	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	75.400%
11) 1,1-Dichloroethene-d2	2.560	65	9330	4.650	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	93.000%
20) 2-Butanone-d5	4.621	46	44868	48.030	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	96.060%
24) Chloroform-d	5.052	84	46338	5.237	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.800%
26) 1,2-Dichloroethane-d4	5.692	65	23246	5.629	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	112.600%
32) Benzene-d6	5.718	84	87078	5.098	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.000%
36) 1,2-Dichloropropane-d6	6.682	67	26958	5.033	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.600%
41) Toluene-d8	7.891	98	71575	4.726	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.600%
43) trans-1,3-Dichloroprop...	8.174	79	9346	5.621	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	112.400%
46) 2-Hexanone-d5	8.627	63	31011	45.222	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	90.440%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	17791	4.705	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.200%
66) 1,2-Dichlorobenzene-d4	12.187	152	24085	4.987	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041824\
Data File : VU058759.D
Acq On : 18 Apr 2024 13:51
Operator : MD/SY
Sample : VIBLK008
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK008

Quant Time: Apr 19 06:55:49 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR040424WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Apr 18 05:04:53 2024
Response via : Initial Calibration

