

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041923\
 Data File : VU053789.D
 Acq On : 19 Apr 2023 18:42
 Operator : JC/MD
 Sample : 02035-01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 38150-121121

Quant Time: Apr 20 13:45:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U041923DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Thu Apr 20 13:45:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.108	96	30790	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.629	95	9511	0.788	ug/l	0.00
Spiked Amount 1.000			Recovery	=	79.000%	
68) 1,2-Dichlorobenzene-d4	12.192	152	12777	0.890	ug/l	0.00
Spiked Amount 1.000			Recovery	=	89.000%	
Target Compounds						
						Qvalue
8) 1,1,2-Trichloro-1,2,2-...	2.578	101	450507	42.391	ug/l	95
24) tert-Butyl Alcohol	3.182	59	62537	71.606	ug/l #	1
25) Methyl tert-Butyl Ether	3.362	73	1063804	43.828	ug/l	93
31) Isopropyl Ether	3.996	45	1194063	42.552	ug/l	90
32) Ethyl-t-butyl ether	4.507	59	1194412	42.077	ug/l	95
33) Tert-Amyl methyl ether	5.944	73	1064237	51.310	ug/l	97
73) n-propylbenzene	10.902	120	244347	20.129	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041923\
Data File : VU053789.D
Acq On : 19 Apr 2023 18:42
Operator : JC/MD
Sample : 02035-01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
38150-121121

Quant Time: Apr 20 13:45:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U041923DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Thu Apr 20 13:45:12 2023
Response via : Initial Calibration

