

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041923\
 Data File : VU053790.D
 Acq On : 19 Apr 2023 19:13
 Operator : JC/MD
 Sample : 02304-01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 38150-111522

Quant Time: Apr 20 13:46:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U041923DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Thu Apr 20 13:45:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.108	96	31259	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.629	95	10036	0.819	ug/l	0.00
Spiked Amount 1.000			Recovery	=	82.000%	
68) 1,2-Dichlorobenzene-d4	12.192	152	12781	0.877	ug/l	0.00
Spiked Amount 1.000			Recovery	=	88.000%	
Target Compounds						
8) 1,1,2-Trichloro-1,2,2-...	2.578	101	362244	33.574	ug/l	95
24) tert-Butyl Alcohol	3.182	59	52467	59.175	ug/l #	1
25) Methyl tert-Butyl Ether	3.362	73	418526	16.984	ug/l	93
31) Isopropyl Ether	3.993	45	1207743	42.394	ug/l	90
32) Ethyl-t-butyl ether	4.507	59	1322975	45.907	ug/l	94
33) Tert-Amyl methyl ether	5.944	73	579995	27.544	ug/l	98
73) n-propylbenzene	10.902	120	117701	9.551	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041923\
Data File : VU053790.D
Acq On : 19 Apr 2023 19:13
Operator : JC/MD
Sample : 02304-01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
38150-111522

Quant Time: Apr 20 13:46:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U041923DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Thu Apr 20 13:45:12 2023
Response via : Initial Calibration

