

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042224\
 Data File : VU058769.D
 Acq On : 22 Apr 2024 16:07
 Operator : MD/SY
 Sample : VU0422WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK180

Quant Time: Apr 23 00:10:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR042224WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Apr 22 23:57:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	96750	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	91933	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	38688	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.589	65	37605	4.941	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	98.800%
7) Chloroethane-d5	1.904	69	27884	4.862	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	97.200%
11) 1,1-Dichloroethene-d2	2.557	65	20029	5.227	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	104.600%
20) 2-Butanone-d5	4.621	46	73193	54.741	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	109.480%
24) Chloroform-d	5.052	84	79414	5.104	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.000%
26) 1,2-Dichloroethane-d4	5.692	65	39747	5.271	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.400%
32) Benzene-d6	5.714	84	152576	5.214	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.200%
36) 1,2-Dichloropropane-d6	6.679	67	46051	5.276	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	105.600%
41) Toluene-d8	7.891	98	131087	5.005	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.000%
43) trans-1,3-Dichloroprop...	8.174	79	15724	4.948	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	99.000%
46) 2-Hexanone-d5	8.627	63	48375	51.374	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	102.740%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	27888	5.246	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	105.000%
66) 1,2-Dichlorobenzene-d4	12.187	152	38124	5.654	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	113.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042224\
Data File : VU058769.D
Acq On : 22 Apr 2024 16:07
Operator : MD/SY
Sample : VU0422WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK180

Quant Time: Apr 23 00:10:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR042224WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Apr 22 23:57:56 2024
Response via : Initial Calibration

