

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042424\
 Data File : VU058792.D
 Acq On : 24 Apr 2024 15:04
 Operator : MD/SY
 Sample : P2154-11
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 XA400

Quant Time: Apr 25 01:43:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM042424.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Apr 25 01:37:46 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.258	114	4138	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.418	117	6215	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.808	152	1995	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.590	65	982	0.482	ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery =	96.000%		
4) 1,2-Dichloroethane-d4	5.685	65	1725	0.496	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery =	100.000%		
7) 1,2-Dichloropropane-d6	6.687	67	1638	0.467	ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery =	94.000%		
8) Toluene-d8	7.887	98	4190	0.472	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery =	94.000%		
10) 1,1,2,2-Tetrachloroeth...	10.745	84	1182	0.489	ug/L	0.00
Spiked Amount 0.500	Range 65 - 120		Recovery =	98.000%		
Target Compounds						
					Qvalue	
9) 1,2-Dibromoethane	8.924	107	3612	1.993	ug/L #	95
12) 1,2,3-Trichloropropane	10.814	75	1459	0.780	ug/L	99
13) 1,2-Dibromo-3-chloropr...	12.990	75	546	1.475	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042424\
Data File : VU058792.D
Acq On : 24 Apr 2024 15:04
Operator : MD/SY
Sample : P2154-11
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
XA400

Quant Time: Apr 25 01:43:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM042424.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Apr 25 01:37:46 2024
Response via : Initial Calibration

