

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042424\
 Data File : VU058786.D
 Acq On : 24 Apr 2024 10:41
 Operator : MD/SY
 Sample : VSTD0.108
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VSTD0.1008

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 04/25/2024
 Supervised By :Semsettin Yesilyurt 04/25/2024

Quant Time: Apr 25 01:25:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM042424.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Apr 25 01:24:31 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.258	114	4132	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.418	117	6068	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.825	152	1888	0.500	ug/L	0.02
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.590	65	186	0.085	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.709	65	314	0.090	ug/L	0.02
7) 1,2-Dichloropropane-d6	6.687	67	305	0.089	ug/L	0.00
8) Toluene-d8	7.903	98	764	0.088	ug/L	0.02
10) 1,1,2,2-Tetrachloroeth...	10.745	84	217	0.092	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.590	62	309	0.116	ug/L #	80
6) Trichloroethene	6.544	95	394	0.105	ug/L	97
9) 1,2-Dibromoethane	8.924	107	192	0.108	ug/L #	89
12) 1,2,3-Trichloropropane	10.814	75	180	0.102	ug/L #	94
13) 1,2-Dibromo-3-chloropr...	13.007	75	39m	0.111	ug/L	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042424\
Data File : VU058786.D
Acq On : 24 Apr 2024 10:41
Operator : MD/SY
Sample : VSTD0.108
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTD0.1008

Manual Integrations
APPROVED

Reviewed By : Mahesh Dadoda 04/25/2024
Supervised By : Semsettin Yesilyurt 04/25/2024

Quant Time: Apr 25 01:25:05 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM042424.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Apr 25 01:24:31 2024
Response via : Initial Calibration

