

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042424\
 Data File : VU058791.D
 Acq On : 24 Apr 2024 13:33
 Operator : MD/SY
 Sample : VU0424WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK003

Quant Time: Apr 25 01:42:54 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM042424.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Apr 25 01:37:46 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.258	114	3983	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.419	117	5856	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.808	152	1822	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.590	65	971	0.495	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	100.000%
4) 1,2-Dichloroethane-d4	5.685	65	1745	0.522	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	104.000%
7) 1,2-Dichloropropane-d6	6.687	67	1664	0.504	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	100.000%
8) Toluene-d8	7.887	98	4168	0.498	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	100.000%
10) 1,1,2,2-Tetrachloroeth...	10.745	84	1179	0.518	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	104.000%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042424\
Data File : VU058791.D
Acq On : 24 Apr 2024 13:33
Operator : MD/SY
Sample : VU0424WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK003

Quant Time: Apr 25 01:42:54 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM042424.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Apr 25 01:37:46 2024
Response via : Initial Calibration

